INVITATION FOR QUOTATION

TEQIP-II/2017/ntst/Shopping/10

03-Nov-2017

Sub: Invitation for Quotations for supply of Goods

Dear Sir,

1. You are invited to submit your most competitive quotation for the following goods with item wise detailed specifications given at Annexure I,

Sr. No	Brief Description	Quantity	Delivery Period(In days)	Place of Delivery	Installation Requirement (if any)
1	-80°C refrigeration	1	45	NIT Sikkim, Ravangla	Yes

- 2. Government of India has received a credit from the International Development Association (IDA) towards the cost of the Technical Education Quality Improvement Programme [TEQIP]-Phase II Project and intends to apply part of the proceeds of this credit to eligible payments under the contract for which this invitation for quotations is issued.
- 3. Quotation,
 - 3.1 The contract shall be for the full quantity as described above.
 - 3.2 Corrections, if any, shall be made by crossing out, initialing, dating and re writing.
 - 3.3 All duties and other levies payable by the supplier under the contract shall be included in the unit price.
 - 3.4 Applicable taxes shall be quoted separately for all items.
 - 3.5 The prices quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.

Strain A

American Spilite

- 3.6 The Prices should be quoted in Indian Rupees only.
- 4. Each bidder shall submit only one quotation.
- 5. Quotation shall remain valid for a period not less than **30** days after the last date of quotation submission.
- 6. Evaluation of Quotations,

The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which

- 6.1 are properly signed; and
- 6.2 Confirm to the terms and conditions, and specifications.
- 7. The Quotations would be evaluated for all items together.
- 8. Award of contract:

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.

- 8.1 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.
- 8.2 The bidder whose bid is accepted will be notified of the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order.
- 9. Payment shall be made in Indian Rupees as follows:

Delivery and Installation - 90% of total cost

Satisfactory Acceptance - 10% of total cost

- 10. All supplied items are under warranty of **24** months from the date of successful acceptance of items.
- 11. You are requested to provide your offer latest by 17:00 hours on 01-Dec-2017.
- 12. Detailed specifications of the items are at Annexure I.
- 13. Training Clause (if any)

3/11/12

- 14. Testing/Installation Clause (if any) Installation required
- 15. Information brochures/ Product catalogue, if any must be accompanied with the quotation clearly indicating the model quoted for.
- 16. Sealed quotation to be submitted/ delivered at the address mentioned below, NIT Sikkim Ravangla Campus, South Sikkim, PIN 737139
- 17. We look forward to receiving your quotation and thank you for your interest in this project.

(Authorized Signatory)

Dr. Achintesh Biswas, Nodal Officer, Procurement,

TEQIP III,

NIT Sikkim

Annexure I

Sr. No	Item Name	Specifications
1	-80°C refrigeration	The instrument cabinet should be upright and single door with approximately 380L capacity
	र विकास में का वास	2. Polyurethane insulation with 5.0" foamed-in-place should be present in the instrument
		3. Two access port (1") for inexpedient probes should be present in the system
	ensergas (1802) saus politica val	4. Dust protection of the instrument condenser by washable and easy removable filter is highly preferred
	Trius hortself vi	5. Single handed instrument operation by ergonomic door handle, lock and key facility and padlock compatible should be present in the instrument
		6. Adjustable temperature range of the instrument should be -75°C to -



86°C.

- 7. Digital display and programmable controller system through Microprocessor control and monitoring system should be present in the instrument
- 8. Audible alarm system should be present due to temperature failure, voltage fluctuation etc.
- 9. Power management system with low voltage Surge protection should be available
- 10. At least four (04) or more inner compartment or selves with door should be present in the instrument for cabinet temperature maintenance.
- 11. Presence of heavy gauge, cold-rolled steel exterior construction with a powder coated paint finish for chipping and rust resistance properties
- 12. Easily accessible vacuum relief port should be present in the instrument after door openings
- 13. The instrument should have dual compressor system of preferably 2× 1 H.P; which should be energy efficient with low level of noise; noise level should be mentioned.
- 14. The instrument should be provided with environment friendly refrigerants
- 15. CE Certification of the instrument is preferred
- 16. Upright freeze racks for cryo-cube box (At least 50%), 24 pieces of cryo-boxes, 2 pair of cryo gloves and 72 hrs of battery backup are desirable
- 17. CO2 or LN2 safety back-up systems for sample protection during power or mechanical failure are highly required
- 18. The instrument should be equipped with all electrical fittings, power cables along with suitable voltage stabilizer for smooth operation as per requirement.

A 31"117


FORMAT FOR QUOTATION SUBMISSION

(In letterhead of the supplier with seal)

	Date:
To:	

SI.	Description of	(with full (Including Ex Factory price, excise duty, packing and		Total Price	Sales tax and other taxes payable		
No.	goods (with full			forwarding, transportation, insurance, other local		In In figures	
	Specifications)			costs incidental to delivery and warranty/ guaranty commitments)		%	(B)
) ·		
			1	Total Cost			

Gross Total Cost (A+B): Rs	
e agree to supply the above goods in accordance with the technical specifications for a total contract price of Rs. $$	— (Amount ir
ures) (Rupees ——————amount in words) within the period specified in the Invitation for Quotations.	



Cympletis

We confirm that the normal commercial warranty/ guarantee of ————— months shall apply to the offered items and we also confirm to	0
agree with terms and conditions as mentioned in the Invitation Letter.	

We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf will engage in bribery.

Signature of Supplier

Name:

Address:

Contact No:

3/11/17